

Certificate of Analysis

REDIPOR PREPARED MEDIA



Part Number 4.0032
Old code 101070
Product Description Tryptone Soya Agar 55mm Contact plate triple vent clear 17ml fill. Flow wrap. Sleeve of 10.
Batch reference 24022832
Work Order Number W2406553
Date of Manufacture 28 February 2024
Expiry Date 28 May 2024
pH (T=7.1 to 7.5) 7.4
Sterility Test Pass
(Absence of growth following 72 hours at 30 - 35°C)

Positive Control organisms to show Productivity Ratio (Inoculum level 10-100 cfu)

Organism	Reference	Criteria	Result
<i>Bacillus subtilis</i>	ATCC 6633	0.7 - 2.0	1.1
<i>Escherichia coli</i>	ATCC 8739	0.7 - 2.0	1.1
<i>Pseudomonas aeruginosa</i>	ATCC 9027	0.7 - 2.0	0.9
<i>Staphylococcus aureus</i>	ATCC 6538	0.7 - 2.0	0.9
<i>Aspergillus brasiliensis</i>	ATCC 16404	0.7 - 2.0	0.9
<i>Candida albicans</i>	ATCC 10231	0.7 - 2.0	0.9
<i>Staphylococcus epidermidis</i>	ATCC 12228	0.7 - 2.0	0.9

Released By: Ela Szokalo

Date: 07 March 2024

All our prepared media products are manufactured and tested at our site in Bicester, Oxfordshire under the control of our ISO9001 : 2015 approved quality management system. The generation of this certificate confirms all sterilisation and performance criteria have been achieved

NOTE: Expiry Date only valid if packs stored unopened at Ambient Room Temperature not exceeding 25°C

Cherwell Laboratories Ltd, 7 & 8 Launton Business Centre, Murdock Road, Bicester, OX26 4XB
Tel: +44 (0)1869 355500, Fax: +44 (0)1869 355545, Website: www.cherwell-labs.co.uk